# DICOM Conformance Statement

Product Name: ACUSON Freestyle
Ultrasound System
Release: VA40
Date: December 2016

www.siemens.com/DICOM

SIEMENS

# 1 CONFORMANCE STATEMENT OVERVIEW

The **ACUSON Freestyle™ Ultrasound System** with DICOM option supports the following DICOM Application Entities:

- Verification
  - o Verification AE
- Transfer
  - o Storage AE
  - o Storage Commitment AE
- Workflow Management
  - o Worklist AE
  - o MPPS AE

**Table 1-1:**
**Network Services**

| SOP Classes | Service Class User (SCU) | Service Class Provider (SCP) |
|---|---|---|
| **VERIFICATION** | | |
| **Verification AE** | | |
| Verification | Yes | Yes |
| **TRANSFER** | | |
| **Storage AE** | | |
| Ultrasound Image Storage | Yes | No |
| Ultrasound Multi-frame Image Storage | Yes | No |
| **Storage Commitment AE** | | |
| Storage Commitment Push Model | Yes | No |
| **WORKFLOW MANAGEMENT** | | |
| **Worklist AE** | | |
| Modality Worklist | Yes | No |
| **MPPS AE** | | |
| MPPS (N-Create, N-Set) | Yes | No |

**Table 1-2:**
**UID Values**

| SOP Class Name | SOP Class UID | Category |
|---|---|---|
| **Verification AE** | | |
| Verification | 1.2.840.10008.1.1 | Verification |
| **Storage AE** | | |
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | Transfer |
| Ultrasound Multi-frame Image Storage | 1.2.840.10008.5.1.4.1.1.3.1 | Transfer |
| **Storage Commitment AE** | | |
| Storage Commitment Push Model | 1.2.840.10008.1.20.1 | Transfer |
| **Worklist AE** | | |
| Modality Worklist | 1.2.840.10008.5.1.4.31 | Workflow Management |
| **MPPS AE** | | |

| SOP Class Name | SOP Class UID | Category |
|---|---|---|
| MPPS (N-Create, N-Set) | 1.2.840.10008.3.1.2.3.3 | Workflow Management |

**Table 1-3:**
**Media Services**

| Media Storage Application Profile | Write Files (FSC or FSU) | Read Files (FSR) |
|---|---|---|
| Ultrasound USB Interchange: Single and Multi-Frame Image Display With Spatial Calibration | | |
| STD-US-SC-MF-USB | Yes | No |

# 2 TABLE OF CONTENTS

1 CONFORMANCE STATEMENT OVERVIEW ..... 2
2 TABLE OF CONTENTS ..... 4
3 INTRODUCTION ..... 8
3.1 Audience ..... 8
3.2 Remarks ..... 8
3.3 Terms and Definitions ..... 9
3.4 Basics of DICOM Communication ..... 11
3.5 Abbreviations ..... 12
3.6 References ..... 12
4 NETWORKING ..... 13
4.1 Implementation Model ..... 13
4.1.1 Application Data Flow ..... 13
4.1.2 Functional Definition of AE's ..... 15
4.1.2.1 Verification AE ..... 15
4.1.2.2 Storage AE ..... 15
4.1.2.3 Storage Commitment AE ..... 15
4.1.2.4 Modality Worklist AE ..... 16
4.1.2.5 Modality Performed Procedure Step AE ..... 16
4.1.3 Sequencing of Real-World Activities ..... 16
4.2 AE Specifications ..... 17
4.2.1 Storage AE ..... 17
4.2.1.1 SOP Classes ..... 17
4.2.1.2 Association Policies ..... 18
4.2.1.3 Association Initiation Policy (Storage SCU) ..... 18
4.2.1.4 Association Acceptance Policy ..... 20
4.2.2 Storage Commitment AE ..... 20
4.2.2.1 SOP Classes ..... 20
4.2.2.2 Association Policies ..... 21
4.2.2.3 Association Initiation Policy (Storage Commitment SCU) ..... 21
4.2.2.4 Association Acceptance Policy ..... 23
4.2.3 Worklist AE ..... 23
4.2.3.1 SOP Classes ..... 23
4.2.3.2 Association Policies ..... 24
4.2.3.3 Association Initiation Policy ..... 24
4.2.3.4 Association Acceptance Policy ..... 29
4.2.4 MPPS AE ..... 30
4.2.4.1 SOP Classes ..... 30
4.2.4.2 Association Policies ..... 30
4.2.4.3 Association Initiation Policy ..... 31
4.2.4.4 Association Acceptance Policy ..... 32
4.3 Network Interfaces ..... 32
4.3.1 Physical Network Interface ..... 32
4.3.2 Additonal Protocols ..... 32
4.3.3 IPV4 and IPv6 Support ..... 32
4.4 Configuration ..... 33

4.4.1 ACUSON Freestyle™ TCP/IP Settings ..... 33
4.4.2 ACUSON Freestyle™ DICOM Settings ..... 33
4.4.3 DICOM Settings for Remote DICOM AEs ..... 33
5 MEDIA INTERCHANGE ..... 33
5.1 Implementation Model ..... 33
5.1.1 Application Data Flow Diagram ..... 33
5.1.2 Functional Definitions of AEs ..... 34
5.1.3 Sequencing of Real-World Activities ..... 34
5.1.4 File Meta Information for Implementation Class and Version ..... 34
5.2 AE Specifications ..... 35
5.2.1 Media Storage AE Specification ..... 35
5.2.2 Implementation Identifying Information ..... 35
5.3 Media Storage Application Profile ..... 35
5.3.1 DICOMDIR Keys ..... 35
5.3.2 Compliance to STD-US-SC-MF-USB ..... 37
6 SUPPORT OF CHARACTER SETS ..... 37
6.1 Character Sets for ACUSON Freestyle™ System ..... 37
7 SECURITY ..... 37
7.1 Security Profiles ..... 37
7.2 Association Level Security ..... 37
7.3 Application Level Security ..... 37
8 ANNEXES ..... 37
8.1 IOD Contents ..... 37
8.1.1 Created SOP Instances ..... 37
8.1.1.1 US Image IOD Attributes ..... 38
8.1.1.2 US Multi-Frame Image IOD Attributes ..... 41
8.1.1.3 MPPS: N-CREATE Attributes ..... 45
8.1.1.4 MPPS: N-SET Attributes ..... 47

# LIST OF TABLES

Table 1-1: Network Services ..... 2
Table 1-2: UID Values ..... 2
Table 1-3: Media Services ..... 3
Table 4-1: SOP Classes for Storage AE ..... 17
Table 4-2: DICOM Application Context ..... 18
Table 4-3: Number of Associations as an Association Initiator for Storage AE ..... 18
Table 4-4: DICOM Implementation Class for Storage AE ..... 18
Table 4-5: Association Initiation Policy for Storage AE ..... 18
Table 4-6: Store Presentation Context ..... 19
Table 4-7: Photometric Interpretation ..... 19
Table 4-8: DICOM Command Response Status Handling Behavior ..... 20
Table 4-9: DICOM Command Communication Failure Behavior ..... 20
Table 4-10: SOP Classes for Storage Commitment AE ..... 20
Table 4-11: DICOM Application Context ..... 21
Table 4-12: Number of Associations as an Association Initiator for Storage Commitment AE ..... 21
Table 4-13: DICOM Implementation Class for Storage Commitment AE ..... 21
Table 4-14: Association Initiation Polcy for Storage Commitment AE ..... 21
Table 4-15: Proposed Presentation Contexts for Storage Commitment AE ..... 22
Table 4-16: DICOM Command Response Status Handling Behavior ..... 22
Table 4-17: DICOM Command Communication Failure Behavior ..... 22
Table 4-18: Proposed Presentation Contexts for Storage Commitment AE ..... 23
Table 4-19: DICOM Command Response Status Handling Behavior ..... 23
Table 4-20: SOP Classes for Worklist AE ..... 23
Table 4-21: DICOM Application Context ..... 24
Table 4-22: Number of Associations as an Association Initiator for Worklist AE ..... 24
Table 4-23: DICOM Implementation Class and Version for Worklist AE ..... 24
Table 4-24: Association Policy for Worklist AE ..... 24
Table 4-25: Proposed Presentation Contexts for Worklist AE ..... 25
Table 4-26: Modality Worklist Matching Key Attributes (Broad Query) ..... 25
Table 4-27: Modality Worklist Matching Key Attributes (Patient Based Query) ..... 26
Table 4-28: Modality Worklist C_FIND_RSP Return Key Attributes ..... 26
Table 4-29: DICOM Command Response Status Handling Behavior ..... 29
Table 4-30: DICOM Command Communication Failure Behavior ..... 29
Table 4-31: SOP Classes for MPPS ..... 30
Table 4-32: DICOM Application Context ..... 30
Table 4-33: Number of Associations as an Association Initiator for MPPS AE ..... 30
Table 4-34: DICOM Implementation Class and Version for MPPS AE ..... 31
Table 4-35: Association Initiation Policy for MPPS AE ..... 31
Table 4-36: Proposed Presentation Contexts for MPPS AE ..... 31
Table 4-37: DICOM Command Response Status Handling Behavior ..... 32
Table 4-38: DICOM Command Communication Failure Behavior ..... 32
Table 5-1: Implementation Class – Media Interchange ..... 34
Table 5-2: Application Profiles, Activities, and Roles for DICOM Exchange Media ..... 35
Table 5-3: DICOM Implementation Class and Version for Media Storage AE ..... 35
Table 5-4: DICOMDIR Keys ..... 36
Table 5-5: STD-US-SC-MF-USB Supported SOP Classes ..... 37

## LIST OF FIGURES

Figure 4-1. Implementation Model..... 14
Figure 4-2. Sequence Diagram for Real-World Activities..... 17
Figure 5-1: Media Application Data Flow Diagram..... 34

# 3 INTRODUCTION

This document describes the conformance to the ACR-NEMA DICOM 3.0 Standard by the ACUSON Freestyle ™ ultrasound system, version VA40 with DICOM option from Siemens Healthineers. It shall establish the conformance specifications for this system only, and does not apply to other products offered by Siemens Healthineers or its affiliates.

The ACUSON Freestyle™ system is a device that generates ultrasound images that can be sent using DICOM standard protocols and definitions to other DICOM compliant devices that support the SOP classes defined in Table 4.1.

This document is written with respect to the adopted portions of the DICOM standard, Version 3. The following sections of this document follow the outline specified in the DICOM Standard NEMA publication PS3.2.[1]

## 3.1 Audience

This document is written for the people that need to understand how the ACUSON Freestyle™ system will integrate into their healthcare facility. This includes both those responsible for overall imaging network policy and architecture, as well as integrators who need to have a detailed understanding of the DICOM features of the product. This document contains some basic DICOM definitions so that any reader may understand how this product implements DICOM features. However, integrators are expected to fully understand all the DICOM terminology, how the tables in this document relate to the product's functionality, and how that functionality integrates with other devices that support compatible DICOM features.

## 3.2 Remarks

The scope of this DICOM Conformance Statement is to facilitate integration between the ACUSON Freestyle™ system and other DICOM products. The Conformance Statement should be read and understood in conjunction with the DICOM Standard.

DICOM, by itself, does not guarantee interoperability. However, the Conformance Statement facilitates a first-level validation for interoperability between different applications supporting the same DICOM functionality as SCU and SCP, respectively.

This Conformance Statement is not intended to replace validation with other DICOM equipment to ensure proper exchange of intended information.

The scope of this Conformance Statement is to facilitate communication with Siemens and other vendors' medical equipment. The Conformance Statement should be read and understood in conjunction with the DICOM 3.0 Standard [1]. However, by itself it is not guaranteed to ensure the desired interoperability and a successful interconnectivity.

The user should be aware of the following important issues:

- The comparison of different conformance statements is the first step towards assessing interconnectivity.
- Test procedures should be defined and executed to validate the required level of interoperability with specific compatible DICOM equipment, as established by the healthcare facility.
- The standard will evolve to meet the users' future requirements. Siemens is actively involved in developing the standard further and therefore reserves the right to make changes to its products or to discontinue its delivery.

[1] Source: DICOM® Standards Publication Part 2, © NEMA. The DICOM Standard is under continuous maintenance. The current official version is available at http://dicom.nema.org.

Siemens reserves the right to modify the design and specifications contained herein without prior notice. Please contact your local Siemens representative for the most recent product information.

## 3.3 Terms and Definitions

Informal definitions are provided for the following terms used in this Conformance Statement. The DICOM Standard is the authoritative source for formal definition of these terms.

**Abstract Syntax** – The information agreed to be exchanged between applications, generally equivalent to a Service/Object Pair (SOP) Class. Examples: Verification SOP Class, Modality Worklist Information Model Find SOP Class, Computed Radiography Image Storage SOP Class.

**ACUSON Freestyle™ system** – The ultrasound system described by this DICOM Conformance Statement.

**Application Entity (AE)** – An end point of a DICOM information exchange, including the DICOM network or media interface software; i.e., the software that sends or receives DICOM information objects or messages. A single device may have multiple Application Entities.

**Application Entity Title** – The externally known name of an *Application Entity*, used to identify a DICOM application to other DICOM applications on the network.

**Application Context** – The specification of the type of communication used between *Application Entities*. Example: DICOM network protocol.

**Association** – A network communication channel set up between *Application Entities*.

**Attribute** – A unit of information in an object definition; a data element identified by a *tag*. The information may be a complex data structure (Sequence), itself composed of lower level data elements. Examples: Patient ID (0010,0020), Accession Number (0008,0050), Photometric Interpretation (0028,0004), Procedure Code Sequence (0008,1032).

**Attribute Macro -** A set of Attributes that are described in a single table that is referenced by multiple Module or other tables.

**Information Object Definition (IOD)** – A data abstraction of a class of similar Real-World Objects which defines the nature and attributes relevant to the class of Real-World objects represented. Examples: US Image IOD, CT Image IOD, Print Job IOD.

**Integrating the Healthcare Enterprise (IHE)** – An initiative sponsored by the Radiological Society of North America (RSNA) to document and demonstrate standards-based methods of sharing information in support of optimal patient care. For additional information, see www.rsna.org/ihe.

**Functional Group -** A set of logically related Attributes that are likely to vary together. May be used in Multi-frame IODs to describe parameters which change on a per frame basis.

**Joint Photographic Experts Group (JPEG) –** Joint Photographic Experts Group, The group was organized in 1986, issuing a standard in 1992, which was approved in 1994 as ISO 10918-1. The JPEG standard is used by DICOM applications.

**Media Application Profile** – The specification of DICOM information objects and encoding exchanged on removable media (e.g., CDs), see DICOM PS3.11.

**Module** – A set of *Attributes* within an *Information Object Definition* that are logically related to each other. Example: Patient Module includes (among others) Patient Name, Patient ID, Patient Birth Date, and Patient Sex.

**Negotiation** – First phase of *Association* establishment that allows *Application Entities* to agree on the types of data to be exchanged and how that data will be encoded.

**Picture Archiving and Communications Systems (PACS)** – A DICOM server that accepts medical images from another DICOM system and stores the images for later retrieval.

**Presentation Context** – The set of DICOM network services used over an *Association*, as negotiated between *Application Entities*; includes *Abstract Syntaxes* and *Transfer Syntaxes*.

**Protocol Data Unit (PDU)** – A packet (piece) of a DICOM message sent across the network. It contains protocol control information and user data. Devices must specify the maximum size packet they can receive for DICOM messages.

**Request (RQ)** – A request from one DICOM AE for service from another DICOM AE.

**Response (RSP)** – A response from one DICOM AE to the request for service from another DICOM AE.

**Security Profile** – A set of mechanisms, such as encryption, user authentication, or digital signatures, used by an *Application Entity* to ensure confidentiality, integrity, and/or availability of exchanged DICOM data.

**Service Class Provider (SCP)** – The role of an *Application Entity* that provides a DICOM network service; typically, a server that performs operations requested by another *Application Entity* (*Service Class User*). Examples: Picture Archiving and Communication System (Image Storage SCP, and image query/retrieve SCP), Radiology Information System (Modality Worklist SCP).

**Service Class User (SCU)** – The role of an *Application Entity* that uses a DICOM network service; typically, a client. Examples: imaging modality (image storage SCU, and modality worklist SCU), imaging workstation (image query/retrieve SCU).

**Service/Object Pair (SOP) Class** – The specification of the network or media transfer (service) of a particular type of data (object); the fundamental unit of DICOM interoperability specification. Examples: Ultrasound Image Storage Service, Basic Grayscale Print Management.

**Service/Object Pair (SOP) Instance** – An information object; a specific occurrence of information exchanged in a *SOP Class*. Examples: a specific x-ray image.

**Tag** – A 32-bit identifier for a data element, represented as a pair of four digit hexadecimal numbers, the “group” and the “element”. If the “group” number is odd, the tag is for a private (manufacturer-specific) data element. Examples: (0010,0020) [Patient ID], (07FE,0010) [Pixel Data], (0019,0210) [private data element].

**Transfer Syntax** – The encoding used for exchange of DICOM information objects and messages. Examples: *JPEG* compressed (images), little endian explicit value representation.

**Unique Identifier (UID)** – A globally unique “dotted decimal” string that identifies a specific object or a class of objects; an ISO-8824 Object Identifier. Examples: Study Instance UID, SOP Class UID, SOP Instance UID.

**Value Representation (VR)** – The format type of an individual DICOM data element, such as text, an integer, a person’s name, or a code. DICOM information objects can be transmitted with either explicit identification of the type of each data element (Explicit VR), or without explicit identification (Implicit VR); with Implicit VR, the receiving application must use a DICOM data dictionary to look up the format of each data element.

## 3.4 Basics of DICOM Communication

This section describes terminology used in this Conformance Statement for the non-specialist. The key terms used in the Conformance Statement are highlighted in *italics* below. This section is not a substitute for training about DICOM, and it makes many simplifications about the meanings of DICOM terms.

Two *Application Entities* (devices) that want to communicate with each other over a network using DICOM protocol must first agree on several things during an initial network “handshake”. One of the two devices must initiate an *Association* (a connection to the other device), and ask if specific services, information, and encoding can be supported by the other device (*Negotiation*).

DICOM specifies a number of network services and types of information objects, each of which is called an *Abstract Syntax* for the Negotiation. DICOM also specifies a variety of methods for encoding data, denoted *Transfer Syntaxes*. The Negotiation allows the initiating Application Entity to propose combinations of Abstract Syntax and Transfer Syntax to be used on the Association; these combinations are called *Presentation Contexts.* The receiving Application Entity accepts the Presentation Contexts it supports.

For each Presentation Context, the Association Negotiation also allows the devices to agree on *Roles* – which one is the *Service Class User* (SCU - client) and which is the *Service Class Provider* (SCP - server). Normally the device initiating the connection is the SCU, i.e., the client system calls the server, but not always.

The Association Negotiation finally enables exchange of maximum network packet (*PDU*) size, security information, and network service options (called *Extended Negotiation* information).

The Application Entities, having negotiated the Association parameters, may now commence exchanging data. Common data exchanges include queries for worklists and lists of stored images, transfer of image objects and analyses (structured reports), and sending images to film printers. Each exchangeable unit of data is formatted by the sender in accordance with the appropriate *Information Object Definition*, and sent using the negotiated Transfer Syntax. There is a Default Transfer Syntax that all systems must accept, but it may not be the most efficient for some use cases. Each transfer is explicitly acknowledged by the receiver with a *Response Status* indicating success, failure, or that query or retrieve operations are still in process.

Two Application Entities may also communicate with each other by exchanging media (such as a USB memory device). Since there is no Association Negotiation possible, they both use a *Media Application Profile* that specifies “pre-negotiated” exchange media format, Abstract Syntax, and Transfer Syntax.

## 3.5 Abbreviations

| | |
|---|---|
| ACR | American College of Radiology |
| AE | DICOM Application Entity |
| AET | DICOM Application Entity Title |
| ASCII | American Standard Code for Information Interchange |
| DB | Database |
| DCS | DICOM Conformance Statement |
| DIMSE | DICOM Message Service Element |
| DSF | Decompressed Single Frame (ACUSON Freestyle™ export option) |
| FSC | Interchange Media File Set Creator |
| FSU | Interchange Media File Set Updater |
| HIS | Hospital Information System |
| ILE | Implicit Little Endian transfer syntax |
| IOD | DICOM Information Object Definition |
| MPPS | Modality Performed Procedure Step |
| MWL | Modality Worklist |
| NEMA | National Electrical Manufacturers Association |
| PACS | Picture Archiving and Communications Systems |
| PDU | DICOM Protocol Data Unit |
| RIS | Radiology Information System |
| RWA | Real-World Activity |
| SC | Storage Commitment |
| SCP | DICOM Service Class Provider (DICOM server) |
| SCU | DICOM Service Class User (DICOM client) |
| SOP | DICOM Service-Object Pair |
| UID | Unique identifier |
| US | Ultrasound |

## 3.6 References

[1] DICOM® Standards Publication, PS 3.1 – PS 3.18, © NEMA. The DICOM Standard is under continuous maintenance. The current official version is available at http://dicom.nema.org.

# 4 NETWORKING

This section describes the ACUSON Freestyle™ networking services.

## 4.1 Implementation Model

ACUSON Freestyle™ users can store images directly on the system's storage device. Images can also be transferred to DICOM workstations and archive servers on a network. Storage Commitment can be used to insure that patient images and data are safely committed. The system is capable of querying a HIS/RIS, using DICOM Basic Worklist Management Service, for a list of scheduled patient procedures. Performed procedure status and other procedure information can be returned to the HIS/RIS using Modality Performed Procedure Step (MPPS).

### 4.1.1 Application Data Flow

Figure 4-1 provides a functional overview of the ACUSON Freestyle™ system's Application Entities (AE). Relationships between user invoked activities (in the circles at the left of the AE) and the associated real-world activities provided by DICOM service providers (in the circles on the right side of the diagram) are shown.




**Figure 4-1. Implementation Model**

## 4.1.2 Functional Definition of AE's

### 4.1.2.1 Verification AE

Verification is available in the DICOM setup page. Verification can be used to send a DICOM verification request (C-ECHO) to a remote Application Entity (AE) and will listen for a response. Verification on the ACUSON Freestyle™ can be performed as an SCU or an SCP.

When the Verification SCU is invoked, the following messages are returned to the user:

- DICOM C-ECHO successful
- Could not establish association
- DICOM C-ECHO timed out

### 4.1.2.2 Storage AE

The ACUSON Freestyle™ acts as a SCU for the C-STORE DICOM network service.

As an SCU, the ACUSON Freestyle™ Storage Application Entity originates associations for transfer of DICOM Ultrasound single and multiframne images.

The system supports automatic and manual storage of saved images. Manual transfers can be initiated through the export dialog, opened from the Study List page. If configured for automatic transfer, the system will transfer images on study close.

The "on study close" method lets the user store studies more accurately in the server. It is a common practice in Ultrasound to delete some images or modify demographic data before the study is closed.

In the event that the ACUSON Freestyle™ is taken off the network as a portable system or when a network failure occurs during a background store, the Storage SCU maintains a queue of failed C-STORE requests. The queue is retried upon power cycling of the ACUSON Freestyle™.

### 4.1.2.3 Storage Commitment AE

The ACUSON Freestyle™ serves as a SCU for the DICOM Storage Commitment service. In order for the ACUSON Freestyle™ to send the Storage Commitment message, at least one Storage Commitment server must be configured in the system and the a Server must be selected in the Storage Commitment Destination combo box on the Export Dialog.

DICOM Storage Commitment is negotiated after all images have been successfully stored to a DICOM Store SCP.

ACUSON Freestyle™ uses the DICOM Storage Commitment Push Model to inform the server when all stores for a study have been completed. The Storage Commitment SCU uses the N-ACTION primitive to make a request to the SCP for safekeeping of a set of SOP instances (eg: Ultrasound images).

The ACUSON Freestyle™ Storage Commitment SCU also performs a role reversal and becomes the SCP for the N-EVENT-REPORT primitives received from the original SCP (now an SCU).

#### 4.1.2.4 Modality Worklist AE

Patient registration can be automated by using the 'Worklist' Real World Activity. Pressing the ‘Patient’ button on the imaging screen displays the patient registration page. Pressing the 'Worklist' button on this page invokes the Worklist Query screen. The ‘Worklist’ button will only be visible if one or more Worklist servers are configured in the system.

The worklist query can be customized by modifying fields on the Worklist Query page. Patient name fields that are partially filled or empty will be treated as though an implicit wildcard was appended at the end of each field. Patient ID and Accession number fields will be exact match only.

Once a Worklist query is initiated, the Query Status field will be updated to indicate “In Progress”. The user has the option to press the Cancel button, which will abort the query operation. If no matches are found, the Query Status field will be updated to indicate that zero responses were received. If one or more matching patients are found, a pick list of patient exams will be displayed to the user. Each of the fields can be sorted in ascending and descending order.

The user will have the option of selecting a patient study or canceling the operation. Selection of a patient from the list will cause all demographic information for that patient to be loaded in the patient registration screen.

#### 4.1.2.5 Modality Performed Procedure Step AE

The ACUSON Freestyle™ Modality Performed Procedure Step (MPPS) sends event transactions that facilitate the transfer of procedure status and billing information from the ultrasound system to the information system. MPPS is triggered at the start of a study and at the end of a study.

The MPPS AE supports both the N-CREATE and N-SET DIMSE Service Elements. The N-CREATE message is sent at the start of a study and the N-SET is sent when a study is closed.

### 4.1.3 Sequencing of Real-World Activities

This section describes the equencing of Real-World Activities performed by the Application Entities using a UML sequence diagram. Real-World Activities are depicted as vertical bars and arrows show the events exchanged between them.




**Figure 4-2. Sequence Diagram for Real-World Activities**

## 4.2 AE Specifications

### 4.2.1 Storage AE

#### 4.2.1.1 SOP Classes

**Table 4-1:**
**SOP Classes for Storage AE**

| SOP Class Name | SOP Class UID | User of Service (SCU) | Provider of Service (SCP) |
|---|---|---|---|
| **Supported Storage SOP Classes** | | | |
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | Yes | No |
| Ultrasound Multi-frame Image Storage | 1.2.840.10008.5.1.4.1.1.3.1 | Yes | No |

#### 4.2.1.2 Association Policies

##### 4.2.1.2.1 General

**Table 4-2:**
**DICOM Application Context**

| | |
|---|---|
| **Application Context Name** | 1.2.840.10008.3.1.1.1 |

##### 4.2.1.2.2 Number Of Associations

**Table 4-3:**
**Number of Associations as an Association Initiator for Storage AE**

| | |
|---|---|
| **Maximum number of simultaneous associations** | 1 |

##### 4.2.1.2.3 Asynchronous Nature

All associations use the default synchronous mode of operation. Asynchronous Operations Window negotiations are not supported on the ACUSON Freestyle™.

##### 4.2.1.2.4 Implementation Identifying Information

**Table 4-4:**
**DICOM Implementation Class for Storage AE**

| | |
|---|---|
| **Implementation Class UID** | 1.3.12.2.1107.5.5.16 |

#### 4.2.1.3 Association Initiation Policy (Storage SCU)

**Table 4-5:**
**Association Initiation Policy for Storage AE**

| Operation or Real-World Activity | Association for |
|---|---|
| Export Study | C-STORE |

#### 4.2.1.3.1 Activity "Export Study"

##### 4.2.1.3.1.1 Description and Sequencing of Activities

Storage of a DICOM object to an external entity is triggered from the study list page when one or more studies are selected and the "Export" button is pressed or upon study close if auto-send is configured. In either case, a C-STORE request is initiated to the selected C-STORE SCP..

If an association to a remote Application Entity could successfully be established, each image will be transferred one after another via the same open association.

Studies that have failed to export are retried when the system is power cycled.

##### 4.2.1.3.1.2 Proposed Presentation Contexts

For all supported images (see SOP Classes in Table 4.1), the following Transfer Syntaxes are supported.

**Table 4-6:**
**Store Presentation Context**

| Abstract Syntax | | Transfer Syntax | |
|---|---|---|---|
| **Name** | **UID** | **Name List** | **UID List** |
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | Implicit VR Little Endian | 1.2.840.10008.1.2 |
| | | JPEG Lossy (Baseline) | 1.2.840.10008.1.2.4.50 |
| Ultrasound Multi-frame Image Storage | 1.2.840.10008.5.1.4.1.1.3.1 | JPEG Lossy (Baseline) | 1.2.840.10008.1.2.4.50 |

##### 4.2.1.3.1.3 Photometric Interpretation

Photometric Interpretation (color mode of the pixel image data) is not a negotiable parameter in DICOM 3.0. The Photometric Interpretation Attribute (0028,0004) is set depending on the transfer syntax and the system configuration.

**Table 4-7:**
**Photometric Interpretation**

| SOP Class | | Transfer Syntax | | Photometric Interpretation |
|---|---|---|---|---|
| **Name** | **UID** | **Name List** | **UID List** | |
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | Implicit VR Little Endian | 1.2.840.10008.1.2 | RGB |
| | | JPEG Lossy (Baseline) | 1.2.840.10008.1.2.4.50 | YBR_FULL_422 |
| Ultrasound Multi-frame Image Storage | 1.2.840.10008.5.1.4.1.1.3.1 | JPEG Lossy (Baseline) | 1.2.840.10008.1.2.4.50 | YBR_FULL_422 |

#### 4.2.1.3.1.4 SOP Specific Conformance to SOP Classes

Refer to section 8.1.1, Created SOP Instances, for a detailed list of attributes.

**Table 4-8:**
**DICOM Command Response Status Handling Behavior**

| Service Status | Further Meaning | Error Code | Behavior |
|---|---|---|---|
| Error | Sending partially or completely failed | Any none null Code | Failure and / or error code reported to user and storage job is cancelled. |
| Success | Image is successfully stored on file system. | 0000 | Success reported to user |

**Table 4-9:**
**DICOM Command Communication Failure Behavior**

| Exception | Behavior |
|---|---|
| Timeout | Failure reported to user |
| Association Aborted | Failure reported to user and the storage job is cancelled. |

### 4.2.1.4 Association Acceptance Policy

Storage AE does not accept association requests.

## 4.2.2 Storage Commitment AE

### 4.2.2.1 SOP Classes

**Table 4-10:**
**SOP Classes for Storage Commitment AE**

| SOP Class Name | SOP Class UID | User of Service (SCU) | Provider of Service (SCP) |
|---|---|---|---|
| **Supported Storage Commitment SOP Classes** | | | |
| Storage Commitment Push Model | 1.2.840.10008.1.20.1 | Yes | No |

#### 4.2.2.2 Association Policies

##### 4.2.2.2.1 General

**Table 4-11:**
**DICOM Application Context**

| | |
|---|---|
| **Application Context Name** | 1.2.840.10008.3.1.1.1 |

##### 4.2.2.2.2 Number of Associations

**Table 4-12:**
**Number of Associations as an Association Initiator for Storage Commitment AE**

| | |
|---|---|
| **Maximum number of simultaneous associations** | 1 |

##### 4.2.2.2.3 Asynchronous Nature

All associations use the default synchronous mode of operation. Asynchronous Operations Window negotiations are not supported on the ACUSON Freestyle™.

##### 4.2.2.2.4 Implementation Identifying Information

**Table 4-13:**
**DICOM Implementation Class for Storage Commitment AE**

| | |
|---|---|
| **Implementation Class UID** | 1.3.12.2.1107.5.5.16 |

#### 4.2.2.3 Association Initiation Policy (Storage Commitment SCU)

**Table 4-14:**
**Association Initiation Polcy for Storage Commitment AE**

| Operation or Real-World Activity | Association for |
|---|---|
| Storage Commitment | N-ACTION-RQ |

##### 4.2.2.3.1 Activity “Send Storage Commitment”

##### 4.2.2.3.1.1 Description and Sequencing of Activities

After sending a study (images) to the C-STORE SCP, the ACUSON Freestyle™ will initiate an association to the selected Storage Commitment SCP. This association will be used to send the N_ACTION_RQ to the SCP.

#### 4.2.2.3.1.2 Proposed Presentation Contexts

**Table 4-15:**
**Proposed Presentation Contexts for Storage Commitment AE**

| Presentation Context Table | | | | | |
|---|---|---|---|---|---|
| Abstract Syntax | | Transfer Syntax | | Role | Extended Negotiation |
| Name | UID | Name List | UID List | | |
| Storage Commitment Push Model | 1.2.840.10008.1.20.1 | Implicit VR Little Endian | 1.2.840.10008.1.2 | SCU | None |

There is no extended negotiation as an SCU.

#### 4.2.2.3.1.3 SOP Specific Conformance to SOP Classes

**Table 4-16:**
**DICOM Command Response Status Handling Behavior**

| Service Status | Further Meaning | Error Code | Behavior |
|---|---|---|---|
| Error | No Retry on Failure | Any none null Code | Failure reported to user |
| Success | Storage Commitment Reply noticed. | 0000 | Success reported to user |

**Table 4-17:**
**DICOM Command Communication Failure Behavior**

| Exception | Behavior |
|---|---|
| Timeout | Failure reported to user |
| Association Aborted | Failure reported to user |

### 4.2.2.3.2 Activity “Receive Storage Commitment Report”

#### 4.2.2.3.2.1 Description and Sequencing of Activities

The ACUSON Freestyle™ accepts the N-EVENT-REPORT-RQ on a separate association from the N-ACTION-RQ.

#### 4.2.2.3.2.2 Proposed Presentation Contexts

**Table 4-18:**
**Proposed Presentation Contexts for Storage Commitment AE**

| Presentation Context Table | | | | | |
|---|---|---|---|---|---|
| **Abstract Syntax** | | **Transfer Syntax** | | **Role** | **Extended Negotiation** |
| **Name** | **UID** | **Name List** | **UID List** | | |
| Storage Commitment Push Model | 1.2.840.10008.1.20.1 | Implicit VR Little Endian | 1.2.840.10008.1.2 | SCP | None |

There is no extended negotiation as an SCP.

#### 4.2.2.3.2.3 SOP Specific Conformance for SOP Classes

**Table 4-19:**
**DICOM Command Response Status Handling Behavior**

| Service Status | Further Meaning | Error Code | Behavior |
|---|---|---|---|
| Error | No retry on failure. | Any none null Code | Failure to commit reported to user, |
| Success | Storage Commitment Reply noticed. | 0000 | Success reported to user. |

### 4.2.2.4 Association Acceptance Policy

The Storage Commitment AE does not accept association requests.

## 4.2.3 Worklist AE

### 4.2.3.1 SOP Classes

This Application Entity provides Standard Conformance to the SOP Classes listed in the table below.

**Table 4-20:**
**SOP Classes for Worklist AE**

| SOP Class Name | SOP Class UID | User of Service (SCU) | Provider of Service (SCP) |
|---|---|---|---|
| **Supported Worklist SOP Classes** | | | |
| Modality Worklist Information Model – FIND | 1.2.840.10008.5.1.4.31 | Yes | No |

#### 4.2.3.2 Association Policies

##### 4.2.3.2.1 General

**Table 4-21:**
**DICOM Application Context**

| | |
|---|---|
| **Application Context Name** | 1.2.840.10008.3.1.1.1 |

##### 4.2.3.2.2 Number of Associations

**Table 4-22:**
**Number of Associations as an Association Initiator for Worklist AE**

| | |
|---|---|
| **Maximum number of simultaneous associations** | 1 |

##### 4.2.3.2.3 Asynchronous Nature

All associations use the default synchronous mode of operation. Asynchronous Operations Window negotiations are not supported on the ACUSON Freestyle™.

##### 4.2.3.2.4 Implementation Identifying Information

**Table 4-23:**
**DICOM Implementation Class and Version for Worklist AE**

| | |
|---|---|
| **Implementation Class UID** | 1.3.12.2.1107.5.5.16 |

#### 4.2.3.3 Association Initiation Policy

**Table 4-24:**
**Association Policy for Worklist AE**

| Operation or Real-World Activity | Association for |
|---|---|
| Modality Worklist Query | C-FIND |

#### 4.2.3.3.1 Activity "Modality Worklist Query"

##### 4.2.3.3.1.1 Description and Sequencing of Activities

The ACUSON Freestyle™ user initiates a Modality Worklist C-FIND Query from the Worklist page by pressing the "Query" button. The user may configure the query by entering information in various fields on the Worklist page.

When the system successfully establishes an association to the remote application entity, the C-FIND-RQ will be sent and the results will be displayed by the system. A preset number of maximum matching results are accepted, at which point, the ACUSON Freestyle™ AE issues a C-FIND-CANCEL request.

##### 4.2.3.3.1.2 Proposed Presentation Contexts

**Table 4-25:**
**Proposed Presentation Contexts for Worklist AE**

| Presentation Context Table | | | | | |
|---|---|---|---|---|---|
| Abstract Syntax | | Transfer Syntax | | Role | Extended Negotiation |
| Name | UID | Name List | UID List | | |
| Modality Worklist-FIND | 1.2.840.10008.5.1.4.31 | Implicit VR Little Endian | 1.2.840.10008.1.2 | SCU | No |

There is no extended negotiation as an SCU.

##### 4.2.3.3.1.3 SOP Specific Conformance to Modality Worklist Service SOP Class

The following table lists the user configurable matching attributes used in the Broad Query of the Modality Worklist (C-FIND).

**Table 4-26:**
**Modality Worklist Matching Key Attributes (Broad Query)**

| RIS Server Configuration | Attribute Name | Tag | Query Value |
|---|---|---|---|
| | Scheduled Procedure Step Sequence | (0040,0100) | |
| Scheduled Station | >Scheduled Station AE Title | (0040,0001) | If selected, set to the AE Title of the ACUSON Freestyle™ system issuing the C-FIND-RQ |
| SPS Start Date Interval | >Scheduled Procedure Step Start Date | (0040,0002) | Selections are: Today, Yesterday, Tomorrow, All |
| Modality | >Modality | (0008,0060) | Selections are: US, All, or user configurable |

The following table lists the user configurable matching attributes used in the Patient Query of the Modality Worklist (C-FIND).

**Table 4-27:**
**Modality Worklist Matching Key Attributes (Patient Based Query)**

| Worklist Data View Field | Attribute Name | Tag | Query Value |
|---|---|---|---|
| Requested Procedure ID | Requested Procedure ID | (0040,1001) | As entered in the "Requested Procedure ID" field. |
| Accession Number | Accession Number | (0008,0050) | As entered in the "Accession Number" field. |
| Patient's Name | Patient's Name | (0010,0010) | As entered in the "Last Name" and / or "First Name" fields. |
| Patient ID | Patient ID | (0010,0020) | As entered in the "Patient ID" field. |
| Scheduled Start Date | Scheduled Procedure Step Sequence<br>>Scheduled Procedure Step Start Date | (0040,0100)<br>>(0040,0002) | As selected in the "Exam Date" field. |
| Modality | Scheduled Procedure Step Sequence<br>>Modality | (0040,0100)<br>>(0008,0060) | As selected or entered in the "Modality" fields. |
| Station AE Title | Scheduled Procedure Step Sequence<br>>Scheduled Station AE Title | (0040,0100)<br>>(0040,0001) | Selected by the "AET Filter" combo box. |

**Table 4-28:**
**Modality Worklist C_FIND_RSP Return Key Attributes**

| Attribute Name | Tag | Return Key Type | Displayable on UI |
|---|---|---|---|
| **SOP Common** | | | |
| Specific Character Set | (0008,0005) | 1C | |
| **Scheduled Procedure Step** | | | |
| Scheduled Procedure Step Sequence | (0040,0100) | 1 | |
| >Scheduled Station AE Title | (0040,0001) | 1 | Yes |
| >Scheduled Procedure Step Start Date | (0040,0002) | 1 | Yes |
| >Scheduled Procedure Step Start Time | (0040,0003) | 1 | Yes |
| >Modality | (0008,0060) | 1 | Yes |
| >Scheduled Performing Physician's Name | (0040,0006) | 2 | |
| >Scheduled Procedure Step Description | (0040,0007) | 1C | Yes |
| >Scheduled Station Name | (0040,0010) | 2 | |
| >Scheduled Procedure Step Location | (0040,0011) | 2 | |
| >Scheduled Protocol Code Sequence | (0040,0008) | 1C | |

| Attribute Name | Tag | Return Key Type | Displayable on UI |
|---|---|---|---|
| >>Code Value | (0008,0100) | 1C | |
| >>Coding Scheme Designator | (0008,0102) | 1C | |
| >>Coding Scheme Version | (0008,0103) | 3 | |
| >>Code Meaning | (0008,0104) | 3 | |
| >Pre-Medication | (0040,0012) | 2C | |
| >Scheduled Procedure Step ID | (0040,0009) | 1 | Yes |
| >Requested Contrast Agent | (0032,1070) | 2C | |
| >Scheduled Procedure Step Status | (0040,0020) | 3 | |
| >Comments on the Scheduled Procedure Step | (0040,0400) | 3 | |
| **Requested Procedure** | | | |
| Requested Procedure ID | (0040,1001) | 1 | Yes |
| Requested Procedure Description | (0032,1060) | 1C | Yes |
| Requested Procedure Code Sequence | (0032,1064) | 1C | |
| >Code Value | (0008,0100) | 1C | |
| >Code Scheme Designator | (0008,0102) | 1C | |
| >Code Scheme Version | (0008,0103) | 3 | |
| >Code Meaning | (0008,0104) | 3 | |
| Study Instance UID | (0020,000D) | 1 | |
| Referenced Study Sequence | (0008,1110) | 2 | |
| >Referenced SOP Class UID | (0008,1150) | 1C | |
| >Referenced SOP Instance UID | (0008,1155) | 1C | |
| Patient Transport Arrangements | (0040,1004) | 2 | |
| Reason for the Requested Procedure | (0040,1002) | 3 | |
| Requested Procedure Priority | (0040,1003) | 2 | |
| Confidentiality Code | (0040,1008) | 3 | |
| Reporting Priority | (0040,1009) | 3 | |
| Names of Intended Recipients of Results | (0040,1010) | 3 | |
| Requested Procedure Comments | (0040,1400) | 3 | Yes |
| **Imaging Service Request** | | | |
| Accession Number | (0008,0050) | 2 | Yes |
| Requesting Physician | (0032,1032) | 2 | Yes |
| Referring Physician's Name | (0008,0090) | 2 | Yes |
| Imaging Service Request Comments | (0040,2400) | 3 | Yes |
| Requesting Service | (0032,1033) | 3 | |
| Issuing Date of Imaging Service Request | (0040,2004) | 3 | |
| Issuing Time of Imaging Service Request | (0040,2005) | 3 | |
| Order Entered By | (0040,2008) | 3 | |

| Attribute Name | Tag | Return Key Type | Displayable on UI |
|---|---|---|---|
| Order Enterer Location | (0040,2009) | 3 | |
| Order Callback Phone Number | (0040,2010) | 3 | |
| Order Placer Issuer and Number | (0040,2016) | 2 | |
| Order Filler Issuer and Number | (0040,2017) | 2 | |
| **Visit Identification** | | | |
| Admission ID | (0038,0010) | 2 | Yes |
| Institution Name | (0008,0080) | 3 | Yes |
| Institution Address | (0008,0081) | 3 | Yes |
| **Visit Status** | | | |
| Current Patient Location | (0038,0300) | 2 | Yes |
| **Visit Relationship** | | | |
| Referenced Patient Sequence | (0008,1120) | 2 | |
| >Referenced SOP Class UID | (0008,1150) | 2 | |
| >Referenced SOP Instance UID | (0008,1155) | 2 | |
| **Visit Admission** | | | |
| Admitting Diagnoses Description | (0008,1080) | 3 | Yes |
| **Patient Identification** | | | |
| Patient's Name | (0010,0010) | 1 | Yes |
| Patient ID | (0010,0020) | 1 | Yes |
| Other Patient IDs | (0010,1000) | 3 | Yes |
| Other Patient Names | (0010,1001) | 3 | Yes |
| Medical Record Locator | (0010,1090) | 3 | Yes |
| **Patient Demographic** | | | |
| Patients Birthdate | (0010,0030) | 2 | Yes |
| Patient's Sex | (0010,0040) | 2 | Yes |
| Patient's Age | (0010,1010) | 3 | Yes |
| Patient's Size | (0010,1020) | 3 | Yes |
| Patient's Weight | (0010,1030) | 2 | Yes |
| Military Rank | (0010,1080) | 3 | Yes |
| Branch Of Service | (0010,1081) | 3 | Yes |
| Confidentiality constraint on patient data | (0040,3001) | 2 | Yes |
| Patient's Address | (0010,1040) | 3 | Yes |
| Patient's Telephone Numbers | (0010,2154) | 3 | Yes |
| Ethnic Group | (0010,2160) | 3 | Yes |
| Patient Comments | (0010,4000) | 3 | Yes |
| **Patient Medical** | | | |
| Special Needs | (0038,0050) | 3 | Yes |
| Patient State | (0038,0500) | 2 | Yes |
| Pregnancy Status | (0010,21C0) | 2 | Yes |
| Medical Alerts | (0010,2000) | 2 | Yes |

| Attribute Name | Tag | Return Key Type | Displayable on UI |
|---|---|---|---|
| Allergies | (0010,2110) | 2 | Yes |
| Smoking Status | (0010,21A0) | 3 | Yes |
| Last Menstrual Date | (0010,21D0) | 3 | Yes |
| Additional Patient History | (0010,21B0) | 3 | Yes |

#### 4.2.3.3.1.4 SOP Specific Conformance to SOP Classes

**Table 4-29:**
**DICOM Command Response Status Handling Behavior**

| Service Status | Further Meaning | Error Code | Behavior |
|---|---|---|---|
| Error | No Modality Worklist results saved | Any none null Code | Failure reported to user |
| Success | Modality Worklist results saved on system | 0000 | Success reported to user |

**Table 4-30:**
**DICOM Command Communication Failure Behavior**

| Exception | Behavior |
|---|---|
| Timeout | Failure reported to user |

### 4.2.3.4 Association Acceptance Policy

The Worklist AE does not accept Association requests.

## 4.2.4 MPPS AE

### 4.2.4.1 SOP Classes

The MPPS AE provides Standard Conformance to the SOP classes listed in the table below.

**Table 4-31:**
**SOP Classes for MPPS**

| SOP Class Name | SOP Class UID | User of Service (SCU) | Provider of Service (SCP) |
|---|---|---|---|
| **Supported MPPS SOP Classes** | | | |
| MPPS (N-Create, N-Set) | 1.2.840.10008.3.1.2.3.3 | Yes | No |

### 4.2.4.2 Association Policies

#### 4.2.4.2.1 General

**Table 4-32:**
**DICOM Application Context**

| | |
|---|---|
| **Application Context Name** | 1.2.840.10008.3.1.1.1 |

#### 4.2.4.2.2 Number of Associations

**Table 4-33:**
**Number of Associations as an Association Initiator for MPPS AE**

| | |
|---|---|
| **Maximum number of simultaneous associations** | 1 |

#### 4.2.4.2.3 Asynchronous Nature

All associations use the default synchronous mode of operation. Asynchronous Operations Window negotiations are not supported on the ACUSON Freestyle™.

#### 4.2.4.2.4 Implementation Identifying Information

**Table 4-34:**
**DICOM Implementation Class and Version for MPPS AE**

| | |
|---|---|
| **Implementation Class UID** | 1.3.12.2.1107.5.5.16 |

### 4.2.4.3 Association Initiation Policy

**Table 4-35:**
**Association Initiation Policy for MPPS AE**

| Operation or Real-World Activity | Association for |
|---|---|
| Study Creation | N-CREATE |
| Study End | N-SET |

#### 4.2.4.3.1 Activity "Send MPPS"

##### 4.2.4.3.1.1 Description and Sequencing of Activities

The ACUSON Freestyle™ supports the DICOM Modality Performed Procedure Step Service as an SCU. The Modality Performed Procedure Step SCU informs the Performed Procedure Step SCP about the procedure performed at the modality using the N-CREATE and N-SET DIMSE services.

Immediately after a new study is created from a Modality Worklist response, the ACUSON Freestyle™ automatically performs an MPPS N-CREATE-RQ operation with a status of IN-PROGRESS for the newly created Performed Procedure Step. When the current patient procedure ends (either with an Study Close, or start of a new study), the ACUSON Freestyle™ will prompt the user to select a completion status of COMPLETE or DISCONTINUED. The MPPS N-SET-RQ is then sent with the selected status.

##### 4.2.4.3.1.2 Proposed Presentation Contexts

**Table 4-36:**
**Proposed Presentation Contexts for MPPS AE**

| Presentation Context Table | | | | | |
|---|---|---|---|---|---|
| **Abstract Syntax** | | **Transfer Syntax** | | **Role** | **Extended Negotiation** |
| **Name** | **UID** | **Name List** | **UID List** | | |
| Modality Performed Procedure Step | 1.2.840.10008.3.1.2.3.3 | Implicit VR Little Endian | 1.2.840.10008.1.2 | SCU | None |

There is no extended negotiation as an SCU.

#### 4.2.4.3.1.3 SOP Specific Conformance to SOP Classes

Refer to sections 8.1.1.3 and 8.1.1.4 for a detailed list of attributes.

If the N-CREATE message is not sent successfully, it will be retried at the time the N-SET is sent.

**Table 4-37:**
**DICOM Command Response Status Handling Behavior**

| Service Status | Further Meaning | Error Code | Behavior |
|---|---|---|---|
| Failure | If N-CREATE-RQ fails, it is retried before N-SET-RQ is sent. If N-SET fails, study must be reopened, then closed. | Any none null Code | Failure reported to user |
| Success | N-CREATE-RQ or N-SET-RQ accepted by MPPS SCP | 0000 | Success logged |

**Table 4-38:**
**DICOM Command Communication Failure Behavior**

| Exception | Behavior |
|---|---|
| Timeout | Failure reported to user |
| Association Aborted | Failure reported to user |

### 4.2.4.4 Association Acceptance Policy

The MPPS AE does not accept association requests

## 4.3 Network Interfaces

### 4.3.1 Physical Network Interface

The ACUSON Freestyle™ is independent from the physical medium over which TCP/IP executes; it inherits this from the OS system upon which it executes.

### 4.3.2 Additonal Protocols

None.

### 4.3.3 IPV4 and IPv6 Support

Only IPv4 is supported.

## 4.4 Configuration

The DICOM settings used by the ACUSON Freestyle™ are configured from the DICOM setup page accessed from the System page. The network settings for both the wired and wireless networks are configured from the Network setup page.

### 4.4.1 ACUSON Freestyle™ TCP/IP Settings

The following TCP/IP Settings can be configured for both the wired and wireless networks used by the ACUSON Freestyle™:

- IP Address
- Subnet Mask
- Gateway
- Auto IP
- Security settings for wireless networks

### 4.4.2 ACUSON Freestyle™ DICOM Settings

The following settings for the ACUSON Freestyle™ DICOM AEs can be configured:

- AE Title
- DICOM port number

### 4.4.3 DICOM Settings for Remote DICOM AEs

Remote DICOM AEs for Storage, Storage Commitment, Modality Worklist and MPPS can be configured with the following settings:

- IP Address
- AE Title
- Port number
- Alias
- Set as default

# 5 MEDIA INTERCHANGE

## 5.1 Implementation Model

### 5.1.1 Application Data Flow Diagram

The ACUSON Freestyle™ provides the functionality to export DICOM instances and DICOMDIR to an external USB file system (memory stick). All SOP classes defined in Table 4-1: SOP Classes for Storage AE are supported for the export functionality.

**Figure 5-1: Media Application Data Flow Diagram**




### 5.1.2 Functional Definitions of AEs

The ACUSON Freestyle™ is capable of creating a new file set on an external USB file system. A DICOMDIR will be created if one does not exist, or will be updated if it exists.

### 5.1.3 Sequencing of Real-World Activities

Storage of a DICOM object to an external USB media is triggered from the study list page when one or more studies are selected and the “Export” button is pressed and the export format is configured for DICOM Media or DICOM Media DSF (decompressed single frame).

### 5.1.4 File Meta Information for Implementation Class and Version

**Table 5-1:**
**Implementation Class – Media Interchange**

| | |
|---|---|
| File Meta Information Version | 0x0001 |
| Implementation Class UID | 1.3.12.2.1107.5.5.16 |

## 5.2 AE Specifications

### 5.2.1 Media Storage AE Specification

The ACUSON Freestyle™ system Media Storage AE provides conformance to the following DICOM SOP Classes as an FSC and FSU:

**Table 5-2:**
**Application Profiles, Activities, and Roles for DICOM Exchange Media**

| Application Profiles Supported | Real World Activity | Role |
|---|---|---|
| STD-US-SC-MF-USB | Export study to USB | FSC/FSU |

### 5.2.2 Implementation Identifying Information

**Table 5-3:**
**DICOM Implementation Class and Version for Media Storage AE**

| | |
|---|---|
| **Implementation Class UID** | 1.3.12.2.1107.5.5.16 |

## 5.3 Media Storage Application Profile

### 5.3.1 DICOMDIR Keys

The DICOMDIR file will contain the following attributes for the levels Patient - Study - Series - Image (valid for all Application profiles described in this section).

**Table 5-4:**
**DICOMDIR Keys**

| Attribute Name | Tag | Type | Notes |
|---|---|---|---|
| **File-Set Identification** | | | |
| File-set ID | (0004,1130) | 2 | Volume label of media |
| Specific Character Set | (0008,0005) | 1C | |
| **Directory Information** | | | |
| Offset of the First Directory Record of the Root Directory Entry | (0004,1200) | 1 | |
| Offset of the Last Directory Record of the Root Directory Entity | (0004,1202) | 1 | |
| File-set Consistency Flag | (0004,1212) | 1 | 0000H |
| Directory Record Sequence | (0004,1220) | 2 | |
| > Offset of the Next Directory Record | (0004,1400) | 1C | |
| > Record In-use flag | (0004,1410) | 1C | FFFFH |
| > Offset of Referenced Lower-Level Directory Entity | (0004,1420) | 1C | |
| > Directory Record Type | (0004,1430) | 1C | PATIENT, STUDY, SERIES, IMAGE |
| > Record Selection Keys | see below | | |
| **Patient Keys** | | | **Directory Record Type PATIENT** |
| Specific Character Set | (0008,0005) | 1C | |
| Patient's Name | (0010,0010) | 2 | |
| Patient ID | (0010,0020) | 1 | |
| **Study Keys** | | | **Directory Record Type STUDY** |
| Specific Character Set | (0008,0005) | 1C | |
| Study Date | (0008,0020) | 1 | |
| Study Time | (0008,0030) | 1 | |
| Accession Number | (0008,0050) | 2 | |
| Study Description | (0008,1030) | 2 | |
| Study Instance UID | (0020,000D) | 1C | |
| Study ID | (0020,0010) | 1 | Set to 1 |
| **Series Keys** | | | **Directory Record Type SERIES** |
| Specific Character Set | (0008,0005) | 1C | |
| Modality | (0008,0060) | 1 | |
| Series Instance UID | (0020,000E) | 1 | |
| Series Number | (0020,0011) | 1 | |
| **Image Keys** | | | **Directory Record Type IMAGE** |
| Specific Character Set | (0008,0005) | 1C | |
| Referenced File ID | (0004,1500) | 1C | Filename on media |
| Referenced SOP Class UID In File | (0004,1510) | 3 | |
| Referenced SOP Instance UID In File | (0004,1511) | 3 | |
| Referenced Transfer Syntax UID In File | (0004,1512) | 3 | |
| Instance Number | (0020,0013) | 1 | |

### 5.3.2 Compliance to STD-US-SC-MF-USB

For media conforming to the STD-US-SC-MF-CDR profiles the following SOP Classes and transfer syntaxes will be supported as an FSC.

**Table 5-5:**
**STD-US-SC-MF-USB Supported SOP Classes**

| IOD | SOP Class UID | Transfer Syntax and UID | FSC | FSR | FSU |
|---|---|---|---|---|---|
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | Explicit VR Little Endian 1.2.840.10008.1.2.1 | Yes | No | No |
| Ultrasound Image Storage | 1.2.840.10008.5.1.4.1.1.6.1 | JPEG Lossy (Baseline) 1.2.840.10008.1.2.4.50 | Yes | No | No |
| Ultrasound Multi-frame Image Storage | 1.2.840.10008.5.1.4.1.1.3.1 | JPEG Lossy (Baseline) 1.2.840.10008.1.2.4.50 | Yes | No | No |

# 6 SUPPORT OF CHARACTER SETS

## 6.1 Character Sets for ACUSON Freestyle™ System

The ACUSON Freestyle™ system supports the ISO 8859 Latin 1 (ISO-IR-100) character set.

# 7 SECURITY

## 7.1 Security Profiles

None Supported.

## 7.2 Association Level Security

None Supported.

## 7.3 Application Level Security

None Supported.

# 8 ANNEXES

## 8.1 IOD Contents

### 8.1.1 Created SOP Instances

#### 8.1.1.1 US Image IOD Attributes

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| Patient Identification | Patient's Name | (0010,0010) | 2 | User input or MWL |
| | Patient ID | (0010,0020) | 2 | User input or MWL |
| | Other Patient Ids | (0010,1000) | 3 | From MWL |
| | Other Patient Names | (0010,1001) | 3 | From MWL |
| | Referenced Patient Sequence | (0008,1120)) | 3 | From MWL |
| Patient Demographic | Patient's Age | (0010,1010) | 2 | From MWL |
| | Confidentiality Constraint on Patient Data Description | (0040,3001) | 3 | From MWL |
| | Patient's Birth Date | (0010,0030) | 2 | User Input or MWL |
| | Patient's Sex | (0010,0040) | 2 | User input or MWL |
| | Patient Size | (0010,1020) | 3 | User input or MWL |
| | Patient Weight | (0010,1030) | 3 | User input or MWL |
| | Patient Address | (0010,1040) | 3 | From MWL |
| | Military Rank | (0010,1080) | 3 | From MWL |
| | Branch Of Service | (0010,1081) | 3 | From MWL |
| | Ethnic Group | (0010,2160) | 3 | From MWL |
| | Patient Comments | (0010,4000) | 3 | From MWL |
| Patient Study | Admitting Diagnosis Description | (0008,1080) | 3 | User input or MWL |
| Patient Medical | Medical Alert | (0010,2000) | 3 | From MWL |
| | Allergies | (0010,2110) | 3 | From MWL |
| | Smoking Status | (0010,21A0) | 3 | From MWL |
| | Additional Patient History | (0010,21B0) | 3 | From MWL |
| | Pregnancy Status | (0010,21C0) | 3 | From MWL |
| | Last Menstrual Date | (0010,21D0) | 3 | From MWL |
| | Special Needs | (0038,0050) | 3 | From MWL |
| | Patient State | (0038,0500) | 3 | From MWL |
| General Study | Study Instance UID | (0020,000D) | 1 | Created or from MWL |
| | Study Date | (0008,0020) | 2 | Date of study creation |
| | Study Time | (0008,0030) | 2 | Time of study creation |
| | Referring Physician's Name | (0008,0090) | 2 | From MWL |
| | Study ID | (0020,0010) | 2 | Created |
| | Accession Number | (0008,0050) | 2 | User input |
| | Study Description | (0008,1030) | 3 | From exam type or "unknown" |
| | Referenced Study Sequence | (0008,1110) | 3 | From MWL |
| | Study Comments | (0032,4000) | 3 | User input |
| General Series | Modality | (0008,0060) | 1 | US |
| | Series Instance UID | (0020,000E) | 1 | Created |
| | Series Number | (0020,0011) | 2 | Created |
| | Performing Physician Name | (0008,1050) | 3 | User input |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | Referenced Performed Procedure Step Sequence | (0008,1111) | 3 | If MPPS is in use |
| | >Referenced SOP Class UID | (0008,1150) | 1 | 1.2.840.10008.3.1.2.3.3 (MPPS SOP Class) |
| | >Referenced SOP Instance UID | (0008,1155) | 1 | MPPS SOP Instance UID |
| | Operator's Name | (0008,1070) | 3 | User input or MWL |
| | Request Attributes Sequence | (0040,0275) | 3 | From MWL or omitted |
| | >Requested Procedure ID | (0040,1001) | 1C | From MWL |
| | >Requested Procedure Description | (0032,1060) | 3 | From MWL |
| | >Requested Procedure Code Sequence | (0032,1064) | 3 | From MWL |
| | >>Code Value | (0008,0100) | 1 | From MWL |
| | >> Coding Scheme Designator | (0008,0102) | 1 | From MWL |
| | >> Coding Scheme Version | (0008,0103) | 1C | From MWL |
| | >> Code Meaning | (0008,0104) | 1 | From MWL |
| | >Scheduled Procedure Step ID | (0040,0009) | 1C | From MWL |
| | >Scheduled Procedure Step Description | (0040,0007) | 3 | From MWL |
| | >Scheduled Procedure Code Sequence | (0040,0008) | 3 | From MWL |
| | >>Code Value | (0008,0100) | 1 | From MWL |
| | >> Coding Scheme Designator | (0008,0102) | 1 | From MWL |
| | >> Coding Scheme Version | (0008,0103) | 1C | From MWL |
| | >> Code Meaning | (0008,0104) | 1 | From MWL |
| Imaging Service Request | Requesting Physician | (0032,1032) | 3 | From MWL |
| | Referring Physician's Name | (0008,0090) | 2 | From MWL |
| | Requesting Service | (0032,1033) | 3 | From MWL |
| | Issue Date of Imaging Request | (0040,2004) | 3 | From MWL |
| | Issue Time of Imaging Request | (0040,2005) | 3 | From MWL |
| | Placer Order Number / Imaging Service Request | (0040,2016) | 3 | From MWL |
| | Filler Order Number / Imaging Service Request | (0040,2017) | 3 | From MWL |
| | Order Entered By | (0040,2008) | 3 | From MWL |
| | Order Enterer Location | (0040,2009) | 3 | From MWL |
| | Order Callback Phone Number | (0040,2010) | 3 | From MWL |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | Admission ID | (0038,0010) | 3 | From MWL |
| | Imaging Service Request Comments | (0040,2400) | 3 | From MWL |
| Requested Procedure | Requested Procedure ID | (0040,1001) | 3 | From MWL |
| | Reason For Requested Procedure | (0040,1002) | 3 | From MWL |
| | Requested Procedure Comments | (0040,1400) | 3 | From MWL |
| | Requested Procedure Code Sequence | (0032,1064) | 3 | From MWL |
| | >Code Sequnce Macro | | | |
| | Requested Procedure Description | (0032,1060) | 3 | From MWL |
| | Requested Procedure Priority | (0040,1003) | 3 | From MWL |
| | Patient Transport Arrangements | (0040,1004) | 3 | From MWL |
| | Confidentiality Code | (0040,1008) | 3 | From MWL |
| | Reporting Priority | (0040,1009) | 3 | From MWL |
| | Names Of Intended Recipients of Results | (0040,1010) | 3 | From MWL |
| Visit Status | Current Patient Location | (0038,0300) | 3 | From MWL |
| General Equipment | Manufacturer | (0008,0070) | 2 | Set to SIEMENS |
| | Institution Name | (0008,0080) | 3 | User input or MWL |
| | Institution Address | (0008,0081) | 3 | From MWL |
| | Manufacturer's Model Name | (0008,1090) | 3 | Set to "ACUSON Freestyle" |
| | Device Serial Number | (0018,1000) | 3 | Set to system serial number |
| | Station Name | (0008,1010) | 3 | Set to system Application Entity Title |
| | Software Version | (0018,1020) | 3 | Set to "VA40x", where x is a letter |
| General Image | Instance Number | (0020,0013) | 2 | Image number in study (1-n) |
| | Patient Orientation | (0020,0020) | 2C | Set to zero length |
| | Content Date | (0008,0023) | 2C | Date of image creation |
| | Content Time | (0008,0033) | 2C | Time of image creation |
| | Image Type | (0008,0008) | 3 | ORIGINAL\PRIMARY |
| | Lossy Image Compression | (0028,2110) | 1C | Set to 1, if JPEG Baseline transfer syntax used, omitted otherwise |
| US Image | Transducer Type | (0018,6031) | 3 | LINEAR or CURVED LINEAR |
| Image Pixel | Samples Per Pixel | (0028,0002) | 1 | Set to 3 |
| | Photometric Interpretation | (0028,0004) | 1 | RGB or YBR_FULL_422 |
| | Rows | (0028,0010) | 1 | Set to 656 |
| | Columns | (0028,0011) | 1 | Set to 800 |
| | Bits Allocated | (0028,0100) | 1 | Set to 8 |
| | Bits Stored | (0028,0101) | 1 | Set to 8 |
| | High Bit | (0028,0102) | 1 | Set to 7 |
| | Pixel Representation | (0028,0103) | 1 | Set to 0 |
| | Pixel Data | (7FE0,0010) | 1 | |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | Planar Configuration | (0028,0006) | 1C | Set to 0 |
| SOP Common | SOP Class UID | (0008,0016) | 1 | Set to 1.2.840.10008.5.1.4.1.1.6.1 |
| | SOP Instance UID | (0008,0018) | 1 | Created |
| | Specific Character Set | (0008,0005) | 1C | |
| US Region Calibration | Sequence of Ultrasound Regions | (0018,6011) | 1 | One for each US region displayed |
| | >Region Location Min x0 | (0018,6018) | 1 | |
| | >Region Location Min y0 | (0018,601A) | 1 | |
| | >Region Location Max x1 | (0018,601C) | 1 | |
| | >Region Location Max y1 | (0018,601E) | 1 | |
| | >Physical Units X Direction | (0018,6024) | 1 | Set to 3 |
| | >Physical Units Y Direction | (0018,6026) | 1 | Set to 3 |
| | >Physical Delta X | (0018,602C) | 1 | |
| | >Physical Delta Y | (0018,602E) | 1 | |
| | >Region Spatial Format | (0018,6012) | 1 | Set to 1 |
| | >Region Data Type | (0018,6014) | 1 | B-Mode set to 1<br>B Mode, with Color set to 2 |
| | >Region Flags | (0018,6016) | 1 | Set to 2 |
| | | | | |
| | | | | |

### 8.1.1.2 US Multi-Frame Image IOD Attributes

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| Patient Identification | Patient's Name | (0010,0010) | 2 | User input or MWL |
| | Patient ID | (0010,0020) | 2 | User input or MWL |
| | Other Patient Ids | (0010,1000) | 3 | From MWL |
| | Other Patient Names | (0010,1001) | 3 | From MWL |
| | Referenced Patient Sequence | (0008,1120)) | 3 | From MWL |
| Patient Demographic | Patient's Age | (0010,1010) | 2 | From MWL |
| | Confidentiality Constraint on Patient Data Description | (0040,3001) | 3 | From MWL |
| | Patient's Birth Date | (0010,0030) | 2 | User Input or MWL |
| | Patient's Sex | (0010,0040) | 2 | User input or MWL |
| | Patient Size | (0010,1020) | 3 | User input or MWL |
| | Patient Weight | (0010,1030) | 3 | User input or MWL |
| | Patient Address | (0010,1040) | 3 | From MWL |
| | Military Rank | (0010,1080) | 3 | From MWL |
| | Branch Of Service | (0010,1081) | 3 | From MWL |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | Ethnic Group | (0010,2160) | 3 | From MWL |
| | Patient Comments | (0010,4000) | 3 | From MWL |
| Patient Study | Admitting Diagnosis Description | (0008,1080) | 3 | User input or MWL |
| Patient Medical | Medical Alert | (0010,2000) | 3 | From MWL |
| | Allergies | (0010,2110) | 3 | From MWL |
| | Smoking Status | (0010,21A0) | 3 | From MWL |
| | Additional Patient History | (0010,21B0) | 3 | From MWL |
| | Pregnancy Status | (0010,21C0) | 3 | From MWL |
| | Last Menstrual Date | (0010,21D0) | 3 | From MWL |
| | Special Needs | (0038,0050) | 3 | From MWL |
| | Patient State | (0038,0500) | 3 | From MWL |
| General Study | Study Instance UID | (0020,000D) | 1 | Created or from MWL |
| | Study Date | (0008,0020) | 2 | Date of study creation |
| | Study Time | (0008,0030) | 2 | Time of study creation |
| | Referring Physician's Name | (0008,0090) | 2 | From MWL |
| | Study ID | (0020,0010) | 2 | Created |
| | Accession Number | (0008,0050) | 2 | User input or MWL |
| | Study Description | (0008,1030) | 3 | From exam type or "unknown" |
| | Referenced Study Sequence | (0008,1110) | 3 | From MWL |
| | Study Comments | (0032,4000) | 3 | User input |
| General Series | Modality | (0008,0060) | 1 | US |
| | Series Instance UID | (0020,000E) | 1 | Created |
| | Series Number | (0020,0011) | 2 | Created |
| | Performing Physician Name | (0008,1050) | 3 | User input |
| | Referenced Performed Procedure Step Sequence | (0008,1111) | 3 | If MPPS is in use |
| | >Referenced SOP Class UID | (0008,1150) | 1 | 1.2.840.10008.3.1.2.3.3 (MPPS SOP Class) |
| | >Referenced SOP Instance UID | (0008,1155) | 1 | MPPS SOP Instance UID |
| | Operator's Name | (0008,1070) | 3 | User input |
| | Request Attributes Sequence | (0040,0275) | 3 | From MWL or omitted |
| | >Requested Procedure ID | (0040,1001) | 1C | From MWL |
| | >Requested Procedure Description | (0032,1060) | 3 | From MWL |
| | >Requested Procedure Code Sequence | (0032,1064) | 3 | From MWL |
| | >>Code Value | (0008,0100) | 1 | From MWL |
| | >> Coding Scheme Designator | (0008,0102) | 1 | From MWL |
| | >> Coding Scheme Version | (0008,0103) | 1C | From MWL |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | >> Code Meaning | (0008,0104) | 1 | From MWL |
| | >Scheduled Procedure Step ID | (0040,0009) | 1C | From MWL |
| | >Scheduled Procedure Step Description | (0040,0007) | 3 | From MWL |
| | >Scheduled Procedure Code Sequence | (0040,0008) | 3 | From MWL |
| | >>Code Value | (0008,0100) | 1 | From MWL |
| | >> Coding Scheme Designator | (0008,0102) | 1 | From MWL |
| | >> Coding Scheme Version | (0008,0103) | 1C | From MWL |
| | >> Code Meaning | (0008,0104) | 1 | From MWL |
| Imaging Service Request | Requesting Physician | (0032,1032) | 3 | From MWL |
| | Referring Physician's Name | (0008,0090) | 2 | From MWL |
| | Requesting Service | (0032,1033) | 3 | From MWL |
| | Issue Date of Imaging Request | (0040,2004) | 3 | From MWL |
| | Issue Time of Imaging Request | (0040,2005) | 3 | From MWL |
| | Placer Order Number / Imaging Service Request | (0040,2016) | 3 | From MWL |
| | Filler Order Number / Imaging Service Request | (0040,2017) | 3 | From MWL |
| | Order Entered By | (0040,2008) | 3 | From MWL |
| | Order Enterer Location | (0040,2009) | 3 | From MWL |
| | Order Callback Phone Number | (0040,2010) | 3 | From MWL |
| | Admission ID | (0038,0010) | 3 | From MWL |
| | Imaging Service Request Comments | (0040,2400) | 3 | From MWL |
| Requested Procedure | Requested Procedure ID | (0040,1001) | 3 | From MWL |
| | Reason For Requested Procedure | (0040,1002) | 3 | From MWL |
| | Requested Procedure Comments | (0040,1400) | 3 | From MWL |
| | Requested Procedure Code Sequence | (0032,1064) | 3 | From MWL |
| | >Code Sequnce Macro | | | |
| | Requested Procedure Description | (0032,1060) | 3 | From MWL |
| | Requested Procedure Priority | (0040,1003) | 3 | From MWL |
| | Patient Transport Arrangements | (0040,1004) | 3 | From MWL |
| | Confidentiality Code | (0040,1008) | 3 | From MWL |
| | Reporting Priority | (0040,1009) | 3 | From MWL |
| | Names Of Intended Recipients of Results | (0040,1010) | 3 | From MWL |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| Visit Status | Current Patient Location | (0038,0300) | 3 | From MWL |
| General Equipment | Manufacturer | (0008,0070) | 2 | Set to SIEMENS |
| | Institution Name | (0008,0080) | 3 | User input or MWL |
| | Institution Address | (0008,0081) | 3 | From MWL |
| | Manufacturer's Model Name | (0008,1090) | 3 | Set to "ACUSON Freestyle" |
| | Device Serial Number | (0018,1000) | 3 | Set to system serial number |
| | Station Name | (0008,1010) | 3 | Set to system Application Entity Title |
| | Software Version | (0018,1020) | 3 | Set to "VA40x", where x is a letter |
| General Image | Instance Number | (0020,0013) | 2 | Image number in study (1-n) |
| | Patient Orientation | (0020,0020) | 2C | Set to zero length |
| | Content Date | (0008,0023) | 2C | Date of image creation |
| | Content Time | (0008,0033) | 2C | Time of image creation |
| | Image Type | (0008,0008) | 3 | ORIGINAL\PRIMARY |
| | Lossy Image Compression | (0028,2110) | 1C | Set to 1, if JPEG Baseline transfer syntax used, omitted otherwise |
| US Image | Transducer Type | (0018,6031) | 3 | LINEAR or CURVED LINEAR |
| Image Pixel | Samples Per Pixel | (0028,0002) | 1 | Set to 3 |
| | Photometric Interpretation | (0028,0004) | 1 | RGB or YBR_FULL_422 |
| | Rows | (0028,0010) | 1 | Set to 656 |
| | Columns | (0028,0011) | 1 | Set to 800 |
| | Bits Allocated | (0028,0100) | 1 | Set to 8 |
| | Bits Stored | (0028,0101) | 1 | Set to 8 |
| | High Bit | (0028,0102) | 1 | Set to 7 |
| | Pixel Representation | (0028,0103) | 1 | Set to 0 |
| | Pixel Data | (7FE0,0010) | 1 | |
| | Planar Configuration | (0028,0006) | 1C | Set to 0 |
| Cine Module | Frame Time | (0018,1063) | 1 | |
| Multi-Frame | Number Of Frames | (0028,0008) | 1 | |
| | Frame Increment Pointer | (0028,0009) | 1 | Set to (0018,1063) |
| SOP Common | SOP Class UID | (0008,0016) | 1 | Set to 1.2.840.10008.5.1.4.1.1.6.1 |
| | SOP Instance UID | (0008,0018) | 1 | Created |
| | Specific Character Set | (0008,0005) | 1C | |
| US Region Calibration | Sequence of Ultrasound Regions | (0018,6011) | 1 | One for each US region displayed |
| | >Region Location Min x0 | (0018,6018) | 1 | |
| | >Region Location Min y0 | (0018,601A) | 1 | |
| | >Region Location Max x1 | (0018,601C) | 1 | |
| | >Region Location Max y1 | (0018,601E) | 1 | |
| | >Physical Units X Direction | (0018,6024) | 1 | Set to 3 |
| | >Physical Units Y Direction | (0018,6026) | 1 | Set to 3 |
| | >Physical Delta X | (0018,602C) | 1 | |

| Module | Attribute | Tag | Type | Notes |
|---|---|---|---|---|
| | >Physical Delta Y | (0018,602E) | 1 | |
| | >Region Spatial Format | (0018,6012) | 1 | Set to 1 |
| | >Region Data Type | (0018,6014) | 1 | B-Mode set to 1<br>B Mode, with Color set to 2 |
| | >Region Flags | (0018,6016) | 1 | Set to 2 |

### 8.1.1.3 MPPS: N-CREATE Attributes

The ACUSON Freestyle™ Performed Procedure Step SCU informs the remote SCU when the scheduled procedure step is started by sending the N-CREATE message. The attributes listed in the table below are sent in the N-CREATE message.

| Attribute Name | Tag | Required Type | Value |
|---|---|---|---|
| **SOP Common** | | | |
| Specific Character Set | (0008,0005) | 1C | From MWL or created |
| **Performed Procedure Step Relationship** | | | |
| Scheduled Step Attribute Sequence | (0040,0270) | 1 | |
| >Study Instance UID | (0020,000D) | 1 | From MWL or created |
| >Referenced Study Sequence | (0008,1110) | 2 | From MWL or zero length |
| >>Referenced SOP Class UID | (0008,1150) | 1C | |
| >>Referenced SOP Instance UID | (0008,1155) | 1C | |
| >Accession Number | (0008,0050) | 2 | From MWL or user input |
| >Requested Procedure ID | (0040,1001) | 2 | From MWL or user input |
| >Requested Procedure Description | (0032,1060) | 2 | From MWL or zero length |
| >Scheduled Procedure Step ID | (0040,0009) | 2 | From MWL or zero length |
| >Scheduled Procedure Step Description | (0040,0007) | 2 | From MWL or zero length |
| >Scheduled Protocol Code Sequence | (0040,0008) | 2 | From MWL or zero length |
| >>Code Value | (0008,0100) | 1C | |
| >>Coding Scheme Designator | (0008,0102) | 1C | |
| >>Code Scheme Version | (0008,0103) | 3 | |
| >>Code Meaning | (0008,0104) | 3 | |
| Patient's Name | (0010,0010) | 2 | From MWL or user input |
| Patient ID | (0010,0020) | 2 | From MWL or user input or created |
| Patients Birth Date | (0010,0030) | 2 | From MWL or user input |
| Patient's Sex | (0010,0040) | 2 | From MWL or user input |
| Referenced Patient Sequence | (0008,1120) | 2 | From MWL or zero length |
| >Referenced SOP Class UID | (0008,1150) | 1C | |
| >Referenced SOP Instance UID | (0008,1155) | 1C | |
| **Performed Procedure Step Information** | | | |
| Performed Procedure Step ID | (0040,0253) | 1 | From SPS ID or created |
| Performed Station AE Title | (0040,0241) | 1 | Own AE Title |
| Performed Station Name | (0040,0242) | 2 | Own hostname |

| Attribute Name | Tag | Required Type | Value |
|---|---|---|---|
| Performed Location | (0040,0243) | 2 | From SPS Location or zero length |
| Performed Procedure Step Start Date | (0040,0244) | 1 | Created |
| Performed Procedure Step Start Time | (0040,0245) | 1 | Created |
| Performed Procedure Step Status | (0040,0252) | 1 | IN PROGRESS |
| Performed Procedure Step Description | (0040,0254) | 2 | From SPS Description or zero length |
| Performed Procedure Type Description | (0040,0255) | 2 | Zero length |
| Procedure Code Sequence | (0008,1032) | 2 | From Requested Procedure Code or zero length |
| >Code Value | (0008,0100) | 1C | |
| >Code Scheme Designator | (0008,0102) | 1C | |
| >Code Scheme Version | (0008,0103) | 3 | |
| >Code Meaning | (0008,0104) | 3 | |
| Performed Procedure Step End Date | (0040,0250) | 2 | Zero length |
| Performed Procedure Step End Time | (0040,0251) | 2 | Zero length |
| **Image Acquisition Results** | | | |
| Modality | (0008,0060) | 1 | US |
| Study ID | (0020,0010) | 2 | From Requested Procedure ID or created |
| Performed Protocol Code Sequence | (0040,0260) | 2 | From Scheduled Action Item Code Sequence or zero length |
| >Code Value | (0008,0100) | 1C | |
| >Code Scheme Designator | (0008,0102) | 1C | |
| >Code Scheme Version | (0008,0103) | 3 | |
| >Code Meaning | (0008,0104) | 3 | |
| Performed Series Sequence | (0040,0340) | 2 | |
| >Performing Physician's Name | (0008,1050) | 2C | Zero Length |
| >Protocol Name | (0018,1030) | 1C | Zero Length |
| >Operator's Name | (0008,1070) | 2C | Zero Length |
| >Series Instance UID | (0020,000E) | 1C | Created |
| >Retrieve AE Title | (0008,0054) | 2C | Zero length |
| >Series Description | (0008,103E) | ?? | |
| >Referenced Image Sequence | (0008,1140) | 2C | Zero length |
| >>Referenced SOP Class UID | (0008,1150) | 1C | |
| >>Referenced SOP Instance UID | (0008,1155) | 1C | |
| >Referenced Non-Image Composite SOP Instance Sequence | (0040,0220) | 2C | Zero Length |

#### 8.1.1.4 MPPS: N-SET Attributes

The ACUSON Freestyle™ Performed Procedure Step SCU informs the remote SCP about the performed procedure step and its status. The N-SET message is only sent when the exam is ended with status “COMPLETED” or when the examination could not be completed with status “DISCONTINUED”. The attributes listed in the table below are sent in the N-SET message.

| Attribute name | Tag | Required Type | Value |
|---|---|---|---|
| **Performed Procedure Step Information** | | | |
| Performed Procedure Step Status | (0040,0252) | 3 | COMPLETED or DISCONTINUED |
| Performed Procedure Step Description | (0040,0254) | 3 | From SPS Description |
| Procedure Code Sequence | (0008,1032) | 3 | From Requested Procedure Code |
| >Code Value | (0008,0100) | 1C | |
| >Code Scheme Designator | (0008,0102) | 1C | |
| >Code Scheme Version | (0008,0103) | 3 | |
| >Code Meaning | (0008,0104) | 3 | |
| Performed Procedure Step End Date | (0040,0250) | 3 | Created |
| Performed Procedure Step End Time | (0040,0251) | 3 | Created |
| **Image Acquisition Results** | | | |
| Performed Action Item Code Sequence | (0040,0260) | 3 | From Scheduled Action Item Code SQ. |
| >Code Value | (0008,0100) | 1C | |
| >Code Scheme Designator | (0008,0102) | 1C | |
| >Code Scheme Version | (0008,0103) | 3 | |
| >Code Meaning | (0008,0104) | 3 | |
| Performed Series Sequence | (0040,0340) | 3 | |
| >Performing Physician’s Name | (0008,1050) | 2C | User input |
| >Protocol Name | (0018,1030) | 1C | Set to “unknown” |
| >Operator’s Name | (0008,1070) | 2C | User input |
| >Series Instance UID | (0020,000E) | 1C | Created |
| >Retrieve AE Title | (0008,0054) | 2C | Zero length |
| >Referenced Image Sequence | (0008,1140) | 2C | Created |
| >>Referenced SOP Class UID | (0008,1150) | 1C | |
| >>Referenced SOP Instance UID | (0008,1155) | 1C | |
| >Referenced Standalone SOP Instance Sequence | (0040,0220) | 2C | Zero length |